


# Meta-analysis of key comparison data in radionuclide metrology


Romain Coulon, Carine Michotte, Vincent Gressier  
International Bureau of Weights and Measures (BIPM)

Presentation at MathMet conference – 2-4 November 2022




# The International System of Reference for radionuclide metrology

- ◆ An ongoing centralised measurement service since 1976



# The International System of Reference for radionuclide metrology (2)

- ◆ A transfer instrument based on ionization chamber measurement




# Key assumptions (1)

Equivalent Activity

$$A_{ei} = A_i \frac{I_{Ra}}{I_i} F_j$$

- ◆ Equivalent activities are assumed to be robust over the long term because:
  - The approach (using a ratio) is intrinsically independent to experimental fluctuations of the efficiency of the ionisation chamber



# Key assumptions (2)

Equivalent Activity


$$A_{ei} = A_i \frac{I_{Ra}}{I_i} F_j$$

## ◆ Other hypothesis are made:

- The homogeneity of batch of glass ampoules with respect to wall thickness
  - ◆ Mitigation: only a unique batch stored by the BIPM is used
- The robustness of the SIR against the choice among the different radium sources
  - ◆ Monitoring: the current ratio between reference sources is monitored



- The robustness of the SIR against the properties of the solution to be measured, notably the filling height and the density.
  - ◆ Mitigation:
    - A range for the mass is specified to participant -  $3.6 \pm 0.2$  g
    - No action with density



# Assessment of the robustness of the SIR


- ◆ The database of the SIR has been updated in a machine readable format permitting to easily implement statistical analysis of the data

[R. Coulon *et al.*, *Meas. Sci. Tech.* **33** 024003, 2022]

- ◆ Knowing the data accumulated since 1976,
  - Is the SIR robust against the choice among the five radium reference sources?



- Is the SIR robust against encountered changes in solution density and volume?



# Data preconditioning

- ◆ The SIR provide one KCRV per radionuclide. To aggregate the data in a common figure of merit, relative degrees of equivalence are used

$$\dot{D}_i^{(rj)} = \frac{D_i^{(rj)}}{KCRV^{(r)}}$$

- ◆ The standard uncertainty of the degrees of equivalence is the combination of the uncertainty from the laboratory  $u\left(A_i^{(r)}\right)$  and the uncertainty from the SIR  $u_{SIR}\left(A_{ei}^{(rj)}\right)$ ,

- ◆ The uncertainty considered in this study is only the SIR component

$$u\left(\dot{D}_i^{(rj)}\right) = \frac{u_{SIR}\left(A_{ei}^{(rj)}\right)}{KCRV^{(r)}}$$

# Data resampling

---

- ◆ To propagate uncertainties through Monte-Carlo method, data set are resampled in a Normal distribution

$$\delta_i^{(rjz)} \sim \mathcal{N} \left( \dot{D}_i^{(rj)}, u^2 \left( u \left( \dot{D}_i^{(rj)} \right) \right) \right)$$

Where,


- $r$  is the radionuclide
- $i$  is the laboratory
- $j$  is radium source number
- $z$  is the data set (MC trial number)



# Model of the data (1)

- ◆ To find the model that best fit the data
  - Parameters maximizing the likelihood of the parametric distributions are estimated
  - Using these parameters, a Kolmogorov-Smirnov (KS) test is applied for each distribution assumptions



Model	MLE p_1	MLE p_2	KS test p-value
Normal	$-3(1) 10^{-4}$	$1.19(8) 10^{-2}$	<b><math>0(3) 10^{-13}</math></b>
Cauchy	$-1.3(6) 10^{-4}$	$3.54(4) 10^{-3}$	<b><math>0.31(6)</math></b>



*Aggregated degrees of equivalence*

# Model of the data (5)

- ◆ Normal hypothesis can also be rejected for the mass and density parameters




- ◆ Marginal distributions of the SIR output, mass and density of the solution are not Normal
- ◆ As a consequence, non parametric tests are implemented

# Correlation with the choice of the reference source (1)

- ◆ To test the absence of correlation with the choice among the 5 reference sources  $j = \{1, 2, 3, 4, 5\}$ , the Kruskal-Wallis and Mood's median tests were implemented.
  - $p$ -values = **0.61(15)** and **0.68(19)** > **0.05**
  - There is no evidence at 95 % confidence that the SIR measurement delivers different expectation values as a function of the reference source used.

*Cauchy-fitted Distributions  
of relative degrees of  
equivalence for the  
subgroups defined by the five  
 $^{226}\text{Ra}$  sources*





# Correlation with the properties of submitted solutions (1)

- ◆ Kendall tau  $\tau$  and Spearman's rank  $r$  correlation coefficients are calculated
  - **$p$ -value of  $\tau_{\delta,m}$  and  $r_{\delta,m} = 0.12(7) > 0.05$**
  - The null hypothesis that **“there is no correlation between the SIR measurement and the mass of submitted solution”** cannot be rejected with a confidence of 95%.
  - This corroborate an initial experimental analysis stating that restraining the volume of solution in the range  $3.6 \pm 0.2$  g should make negligible the impact of the source height on SIR measurements.

[A. Rytz, *Environment International* **1** 15, 1978]

- **$p$ -value of  $\tau_{\delta,\rho} = 0.078(8)$  and  $p$ -value of  $r_{\delta,\rho} = 0.10(1) > 0.05$**
- The null hypothesis that **“there is no correlation between the SIR measurement and the density of submitted solution”** cannot be rejected with a confidence of 95%.
- The impact of the density is strongly radionuclide-dependent because correlated with the energy of x rays and  $\gamma$  rays
- Some experimental studies have underlined the significant impact of the solution density with ionization chamber measurement when the emitted radiation are with low energy – notably for  $^{241}\text{Am}$  (Michotte, for the SIR) and  $^{57}\text{Co}$  (Cessna, for other types of ICs)


# Correlation with the density of submitted solutions



- ◆ A significant correlation between the SIR measurement and solution density has been revealed for the  $^{109}\text{Cd}$  (88 keV  $\gamma$ -rays and x-rays below 26 keV)

# Conclusion and perspectives

- ◆ A meta-analysis was conducted using the new key comparison database
  - Kolmogorov-Smirnov tests concluded
    - ◆ The model of the measurement data is not normally distributed but rather Cauchy distributed
  - Non-parametric null hypothesis tests concluded
    - ◆ The absence of correlation between the SIR measurement and the reference  $^{226}\text{Ra}$  source used
    - ◆ The absence of correlation between the SIR measurement and the mass of submitted standard solution
    - ◆ The absence of correlation between the SIR measurement and the density of submitted standard solution **except for low energy ones emitting only low energy x/ $\gamma$ -rays (e.g.  $^{109}\text{Cd}$ )**
- ◆ The **extension of the SIR** based on liquid scintillation counting will improve the robustness of BIPM services **for radionuclides where the SIR cannot measure or is proven not to being robust.**




# Back up

---

# Model of the data (2)

- ◆ When looking at the data grouped per radium sources




The Normal hypothesis can be rejected with a confidence level of 95% for all the radium sources except the number 5

The Cauchy distribution better fit the data



# Model of the data (3) - $p(D_i, r)$

- When looking at the data grouped per radionuclides



The Normal hypothesis can be rejected for  $^{133}\text{Ba}$ ,  $^{57}\text{Co}$ ,  $^{60}\text{Co}$ ,  $^{59}\text{Fe}$ ,  $^{88}\text{Y}$  and  $^{65}\text{Zn}$

# Model of the data (4)

- ◆ The deviation for the Normal hypothesis can come from
  - The aggregation of SIR measurements (independent measurements - outliers)
  - The SIR measurement model

$$? \longrightarrow A_{ei} = A_i \frac{I_{Ra}}{I_i}$$

Normal r.v. (pointing to  $I_{Ra}$ )  
Normal r.v. (pointing to  $I_i$ )  
Normal r.v. (pointing to the ratio)

- We tend to observe a compliance with the normal hypothesis when:
  - ◆  $u(I_i) \ll u(A_i)$
- Indeed, the ratio of normally distributed r.v. is Cauchy distributed
- So, the normality of the SIR output value cannot be systematically assumed